

Parmax AB
Finspångsgatan 42
SE-163 53 Spånga
Sweden
2023-06-20

Notified Body Confirmation Letter
Reference: 41319273-02 – CN00288-01

To whom it may concern,

Confirmation of the status of a formal application, written agreement, and appropriate surveillance in the framework of Regulation EU 2023/607 amending Regulations (EU) 2017/745 and (EU) 2017/746 as regards the transitional provisions for certain medical devices and in vitro diagnostic medical devices

This letter confirms that, Intertek Medical Notified Body AB, a Notified Body (NB) designated against Regulation (EU) 2017/745 (MDR) and identified by the number 2862 on NANDO, has received a formal application in accordance with Section 4.3, first subparagraph of Annex VII of MDR and has signed a written agreement in accordance with Section 4.3, second subparagraph of Annex VII of MDR with the following manufacturer:

Parmax AB
Finspångsgatan 42
SE-163 53 Spånga
Sweden

The devices covered by the formal application and the written agreement mentioned above are identified in the Tables below. Table 1 identifies the devices for which an MDR application has been received, written agreement concluded and for which the NB is also responsible for appropriate surveillance of the corresponding devices under the applicable Directive. Table 2 identifies the devices for which an MDR application has been received and a written agreement concluded, but the NB has not yet taken the responsibility for appropriate surveillance of the corresponding devices under the applicable Directive.

In the case of devices covered by certificates issued under Directive 90/385/EEC (AIMDD) or Directive 93/42/EEC (MDD) that expired after 26 May 2021 and before 20 March 2023, without having been withdrawn, this letter also confirms that the manufacturer signed the written agreement under MDR by the date of MDD/AIMDD certificate expiry; or provided evidence that a competent authority of a Member State had granted a derogation or exemption from the applicable conformity assessment procedure in accordance with Article 59(1) of MDR or Article 97(1) of the MDR respectively, by the 20 Mar 2023 for the relevant devices.

The transition timelines that apply to the devices covered by this letter, subject to the manufacturer's continued compliance to the other conditions specified in Article 120.3c of MDR (as amended by (EU) 2023/607), are shown below:

- 26 May 2026 for Class III custom-made implantable devices
- 31 December 2027 for Class III devices and Class IIb implantable devices excluding Well-established technologies (WET - sutures, staples, dental fillings, dental braces, tooth crowns, screws, wedges, plates, wires, pins, clips and connectors)
- 31 December 2028 for other Class IIb devices, Class IIa, Class I devices placed on the market in sterile condition or have a measuring function
- 31 December 2028 for devices not requiring the involvement of a notified body under MDD but requiring it under MDR (e.g., class I devices that qualify as re-usable surgical instruments)

On behalf of the Notified Body,



Brian Mather
Certification Manager
Intertek Medical Notified Body AB

Table 1: Devices covered by this letter and for which the NB is also responsible for appropriate surveillance of the corresponding devices under the applicable Directive:

Ref Number/ Device Identification	Device Name	Device classification	MDD Certificate Reference(s)
PT-XY S	Parmax Classic Titanium, 7.8 mm	Ila	41319273-02
PT-XY M	Parmax Classic Titanium, 9.3 mm	Ila	41319273-02
PT-XY L	Parmax Classic Titanium, 11.8 mm	Ila	41319273-02
PT-XY EL	Parmax Classic Titanium, 14.2 mm	Ila	41319273-02
PT-XY SL	Parmax Classic Titanium, 17 mm	Ila	41319273-02
PG-XY	Parmax Classic Gold Plated,	Ila	41319273-02
PG-XY	Paralight Small	Ila	41319273-02
PG-XY	Paralight Medium	Ila	41319273-02
PG-XY	Paralight Large	Ila	41319273-02
PT-60	Paramax Classic Titanium (60 pieces assorted kit)	Ila	41319273-02
PT-120	Paramax Classic Titanium (120 pieces assorted kit)	Ila	41319273-02
PG-60	Paramax Classic Gold (60 pieces assorted kit)	Ila	41319273-02
PG-120	Paramax Classic Gold (120 pieces assorted kit)	Ila	41319273-02
PG-240	Paramax Classic Gold (240 pieces assorted kit)	Ila	41319273-02
PRX-Y A	Parmax Reamers, (short) 28 mm	Ila	41319273-02
PRX-Y B	Parmax Reamers, (long) 33 mm	Ila	41319273-02
PLR-z	Parmax Paralight Reamers, , z = Size, #3 x-small, short (28 mm) #4 small, long (33 mm), #5 medium , long (33 mm), #6 large, , long (33 mm)	Ila	41319273-02
PRA-1/6	6 pack, short reamers size 1-6	Ila	41319273-02
PRB-1/6	6 pack, long reamers size 1-6	Ila	41319273-02
PLR-4/6	3-pack , long reamers size 4-6	Ila	41319273-02

Table 2: Devices covered by this letter and for which the NB is NOT responsible for appropriate surveillance of the corresponding devices under the applicable Directive:

Ref Number/ Device Identification	Device Name	Device classification	MDD Certificate Reference(s)
N/A			

Confirmation Letter Revision History

Date	NB internal reference traceable to each version of the letter	Action